

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 29 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130404

Name BRAD REED ID# 6158 Date 7-12-2009

A Agency FORT GEELY Police Dept Phone # 873-9149

Instrument Location Booking Room

B Alco S/N X301572 Target Value .079 High Pressure 300 lbs

Alco Test Values .079 .079
1st Alco 2nd Alco

Signature [Handwritten Signature]

[Handwritten Signature]
7/30/09

(OVER)


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of Aug, 2009.

[Handwritten Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

JULY 12, 2009

OPERATOR'S NAME:

REED, BRAD

OPERATOR'S NUMBER: 6158

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0000000

DEPT/AGENCY: B164

CASE/REPORT: 000000

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: X301572

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 28.95 in
ALCO TARGET .076 19:37
BLANK TEST .000 19:38
INTERNAL STANDARD VERIFIED 19:38
ALCO TV 28.95 in .079 19:38
BLANK TEST .000 19:39
SUBJECT SAMPLE .000 19:40
BLANK TEST .000 19:41
ALCO TV 28.95 in .079 19:41
BLANK TEST .000 19:42

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

JULY 12, 2009

TIME 19:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c
BAROMETER: 28.95 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~