VERIFICATION OF CALIBRATION REPORT

OCT 0 6 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

Supervisor Operator Performing the Verification Procedure:

A

Name William A. Crabtree ID# 6316 Date 01 Oct 09

Agency Fort Greely Police Department Phone # 873-9120

Instrument Location Booking Room, Bldg 663 Basement

B

Alco S/N X301572 Target Value .079 High Pressure 1000

Alco Test Values .080 (1st Alco) .079 (2nd Alco)

Signature [signature]

(OVER)

10/13/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[signature]

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct., 2009.

[signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

OCTOBER 01, 2009 BMB 10/13/09

OPERATOR'S NAME:
CRABTREE/WILLIAM/A
OPERATOR'S NUMBER: 6316
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 123456
DEPT/AGENCY: BI64
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X301572

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.10 in

ALCO TARGET	.076	23:02
BLANK TEST	.000	23:03
INTERNAL STANDARD	VERIFIED	23:03
ALCO TV 29.10 in	.000	23:03
BLANK TEST	.000	23:04
SUBJECT SAMPLE	.000	23:05
BLANK TEST	.000	23:06
ALCO TV 29.12 in	.079	23:06
BLANK TEST	.000	23:07

BMB 10/13/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

OCTOBER 01, 2009
TIME 22:58

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.12 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~