

JAN 29 2010 ^{CSK 21:40}

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

FEB 01 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130404
A	Name	John W Carroll	ID# 6775 Date 01/13/2010
	Agency	Fort Greely Police BIG 4	Phone# (907) 876-9134
	Instrument Location	Ft. Greely Police Booking Room, Bldg 663	
B	Alco S/N	X172958	Target Value 0.087 High Pressure 1000
	Alco Test Values	0.084	0.084
		1 st Alco	2 nd Alco
	Signature	<i>John W Carroll</i>	
			CSK 2/3/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 10 day of Feb, 2010.

Carolyn M. Noland
 (Notary Seal Stamp)

Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

JANUARY 13, 2010

OPERATOR'S NAME:

CARROLL, JOHN

OPERATOR'S NUMBER: 6775

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: FT GREELY PD B164

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .087

ALCO S/N: X172958

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 28.35 in
ALCO TARGET .082 16:16
BLANK TEST .000 16:17
INTERNAL STANDARD VERIFIED 16:17
ALCO TV 28.35 in .084 16:17
BLANK TEST .000 16:18
SUBJECT SAMPLE .000 16:18
BLANK TEST .000 16:19
ALCO TV 28.35 in .084 16:19
BLANK TEST .000 16:20

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

JANUARY 13, 2010

TIME 16:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 28.35 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { } ~