

VERIFICATION OF CALIBRATION REPORT

of DataMaster edm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster edm S/N <u>130404</u>
Name	<u>John W. Carrall</u>	ID# <u>6775</u> Date <u>May 8, 2010</u>
Agency	<u>Ft. Greely Police Dept. BI64</u>	Phone # <u>907-873-9134</u>
Instrument Location	<u>Bldg 643 Booking Room Ft. Greely PD.</u>	
Alco S/N <u>X172958</u>		Target Value <u>.087</u> High Pressure <u>800</u>
Alco Test Values		
	<u>.087</u> 1 st Alco	<u>.087</u> 2 nd Alco
Signature	<u>John W Carrall</u>	<u>CWB</u> <u>3/24/10</u>
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of June, 2010.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 03:00



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

MAY 08, 2010

OPERATOR'S NAME:

CARROLL, JOHN W

OPERATOR'S NUMBER: 6775

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 000000

DEPT/AGENCY: FT GREELY PD BIG4

CASE/REPORT: 10-VOC

TEST TYPE: V

ALCO TARGET VALUE: .087

ALCO S/N: X172958

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.02 in
ALCO TARGET .084 10:00
BLANK TEST .000 10:01
INTERNAL STANDARD VERIFIED 10:01
ALCO TV 29.02 in .087 10:02
BLANK TEST .000 10:02
SUBJECT SAMPLE .000 10:03
BLANK TEST .000 10:04
ALCO TV 29.02 in .087 10:04
BLANK TEST .000 10:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

MAY 08, 2010

TIME 09:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 40c
BAROMETER: 29.02 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~