VERIFICATION OF CALIBRATION REPORT

SEP 0 3 2010

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

Supervisor/Operator Performing the Verification Procedure:

A

Name Crabtree, William A. ID# 6316 Date 31 Aug 10

Agency Fort Greely Police Department (BIG4) Phone # (907) 873-9120

Instrument Location Bldg 663, Fort Greely PD Booking Room

B

Alco S/N X172958 Target Value .087 High Pressure 800

Alco Test Values .085 (1st Alco) .085 (2nd Alco)

Signature

(OVER)

9/3/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of Oct, 2010.

(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires With Office STATE OF ALASKA

BTP 03-06

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

AUGUST 31, 2010

BMB 9/3/10

OPERATOR'S NAME:
CRABTREE/WILLIAM/A
OPERATOR'S NUMBER: 6316
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
D.L. #: 123456
DEPT/AGENCY: BIG4
CASE/REPORT: 123456
TEST TYPE: V
ALCO TARGET VALUE: .087
ALCO S/N: X172958

— BREATH ANALYSIS —

.087 ADJUSTED FOR 28.60 in

ALCO TARGET	.083	06:05
BLANK TEST	.000	06:06
INTERNAL STANDARD	VERIFIED	06:06
ALCO TV 28.60 in	.085	06:06
BLANK TEST	.000	06:07
SUBJECT SAMPLE	.000	06:07
BLANK TEST	.000	06:08
ALCO TV 28.60 in	.085	06:09
BLANK TEST	.000	06:09

BMB 9/3/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

AUGUST 31, 2010
TIME 06:02

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	28.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

819 03 06