

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

JANUARY 21, 2011

OPERATOR'S NAME: *BMB 3/6/11*
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
V
O.L.L. #: V
DEPT/AGENCY: TAL0
CASE/REPORT: V
TEST TYPE: V
ALCO TARGET VALUE: .002
ALCO S/N: X301501

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.18 in
ALCO TARGET .079 11:15
BLANK TEST .000 11:17
INTERNAL STANDARD VERIFIED 11:17
ALCO TV 29.18 in .002 11:17
BLANK TEST .000 11:18
SUBJECT SAMPLE *BMB 3/6/11* .000 11:18
BLANK TEST .000 11:19
ALCO TV 29.18 in .001 11:19
BLANK TEST .000 11:20

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

JANUARY 21, 2011
TIME 11:21

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/81
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.18 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~