

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

JUL 27 2009


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130492
Name		Shayne Calt	ID# 4559 Date 7-18-09
A	Agency	AST	Phone # 733-2256
Instrument Location		Willow AST	✓
Alco S/N		X301581	Target Value .082 High Pressure 1000
B	Alco Test Value Average	.083	.083
		1 st Alco	2 nd Alco
Signature		A	
		CBS 7/30/09	


(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Aug, 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130492 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

JULY 18, 2009 ✓ *CS 7/20/09*

OPERATOR'S NAME:
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
V
O.L. #: V
DEPT/AGENCY: TALO
CASE/REPORT: V
TEST TYPE: V
ALCO TARGET VALUE: .082 ✓
ALCO S/N: X301581 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.83 in
ALCO TARGET .081 03:01
BLANK TEST .000 03:02
INTERNAL STANDARD VERIFIED 03:02
ALCO TV 29.83 in .083 ✓ 03:03
BLANK TEST .000 03:04
SUBJECT SAMPLE *CS 7/20/09* .000 03:04
BLANK TEST .000 03:05
ALCO TV 29.83 in .083 ✓ 03:05
BLANK TEST .000 03:06

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

JULY 18, 2009 ✓
TIME 02:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.83 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~