


VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N	130492 ✓			
Name	Shayne Calt	ID#	4559	Date	9-18-08 ✓	
A	Agency	AST	Phone #	733-2256		
Instrument Location Willow AST ✓						
Alco S/N		75470 ✓	Target Value	.076 ✓	High Pressure	800
B		Alco Test Value Average	.076 ✓	.075 ✓		
		1 st Alco		2 nd Alco		
Signature					008 9/22/08	
(OVER)						


Under the Alaska Rules of Evidence, I certify that;


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Shayne M Calt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/5/08
Date

SEP 22 2008

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130492

SEPTEMBER 18, 2008

OPERATOR'S NAME:
 CALT SHAYNE H
 OPERATOR'S NUMBER: 4559
 SUBJECT'S LAST NAME:
 VERIFICATION OF CAL
 SUBJECT'S FIRST NAME/MI :
 V
 O.L. #: V
 DEPT/AGENCY: TALQ
 CASE/REPORT: V
 TEST TYPE: V
 ALCO TARGET VALUE: .076
 ALCO S/N: 75478

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.41 in		
ALCO TARGET	.074	18:26
BLANK TEST	.000	18:27
INTERNAL STANDARD	VERIFIED	18:27
ALCO TV 29.41 in	.076	18:27
BLANK TEST	.000	18:28
SUBJECT SAMPLE	.000	18:28
BLANK TEST	.000	18:29
ALCO TV 29.41 in	.075	18:29
BLANK TEST	.000	18:30

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130492

SEPTEMBER 18, 2008
 TIME 18:23

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 02/20/01

HEATERS
 SAMPLE CHAMBER: 49c
 BREATH TUBE: 42c

BAROMETER: 29.41 in

FLOW DETECTOR: OKAY

PUMP
 HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

--- PRINTER TEST ---

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
 H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` { | } ~ ¡ ¢ £ ¤ ¥ ¦ § ¨ © ª « ¬ ® ¯ ° ± ² ³ ´ µ ¶ · ¸ ¹ º » ¼ ½ ¾
                 ¡ ¢ ¤ ¥ ¦ § ¨ © ª « ¬ ­ ¯ ° ± ² ³ ´ µ ¶ · ¸ ¹ º » ¼ ½ ¾ ¿