VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 29 2009

DataMaster cdm S/N 130492

Supervisor/Operator Performing the Verification Procedure:

A

Name Shayne Calt ID# 4559 Date 9-25-09

Agency AST Phone # 733-2258

Instrument Location Willow AST

B

Alco S/N X301581 Target Value .082 High Pressure 1000

Alco Test Value Average .083 (1st Alco) .082 (2nd Alco)

Signature

(OVER)

(Scientific Crime Detection Laboratory)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of Oct., 2009.

(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

1000 psi

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

SEPTEMBER 25, 2009 BMB 10/2/09

OPERATOR'S NAME:
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATIONOFCAL
SUBJECT'S FIRST NAME/MI :
V
O.L. #: V
DEPT/AGENCY: TAL0
CASE/REPORT: V
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: X301581

— BREATH ANALYSIS —

.082 ADJUSTED FOR 29.34 in

ALCO TARGET	.080	18:48
BLANK TEST	.000	18:48
INTERNAL STANDARD	VERIFIED	18:48
ALCO TV 29.34 in	.083	18:49
BLANK TEST	.000	18:50
SUBJECT SAMPLE	.000	18:50
BLANK TEST	.000	18:51
ALCO TV 29.34 in	.082	18:51
BLANK TEST	.000	18:52

BMB 10/2/09

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

SEPTEMBER 25, 2009
TIME 18:45

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.34 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~