

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | |
|--|---------------------------------------|
| DataMaster cdm S/N <u>130493</u> | |
| Supervisor/Operator Performing the Verification Procedure: | |
| Name <u>PATRICK SMITH</u> | ID# <u>6397</u> Date <u>10 APR 11</u> |
| A Agency <u>CLEAR SECURITY FOR LIFE</u> | Phone # _____ |
| Instrument Location <u>BLDG 37, CLEAR AFS</u> | |
| B Alco S/N <u>X124609</u> Target Value <u>.081</u> High Pressure <u>1100</u> | |
| Alco Test Values <u>.076</u> <u>.077</u> | |
| 1st Alco 2nd Alco | |
| Signature <u>[Signature]</u> | |
| (OVER) | |

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2010 2011

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas KYLE KLOPPER
Notary Public, State of Alaska
Commission Expires with Office



CRF
4/21/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

APRIL 12, 2011

CRF 4/21/11

OPERATOR'S NAME:
SMITH/PATRICK/J
OPERATOR'S NUMBER: 6397
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567
DEPT/AGENCY: CLR4
CASE/REPORT: 0987654
TEST TYPE: V
ALCO TARGET VALUE: .001
ALCO S/N: X124609

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.65 in
ALCO TARGET .000 11:21
BLANK TEST .000 11:21
INTERNAL STANDARD VERIFIED 11:22
ALCO TV 29.65 in .076 11:22
BLANK TEST .000 11:23
SUBJECT SAMPLE .000 11:23 CRF 4/21/11
BLANK TEST .000 11:24
ALCO TV 29.65 in .077 11:24
BLANK TEST .000 11:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

APRIL 12, 2011
TIME 11:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.65 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~"

