

JUN 23 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

		DataMaster cdm S/N	130493
Supervisor/Operator Performing the Verification Procedure:			
Name	PATRICK J SMITH	ID #	6397
Date:	Jun 12, 2011		
Agency	CLEAR AFS SECURITY FORCES	Phone #	907-585-6313
Instrument Location	BLDG 37 CLEAR AFS, AK		
Alco S/N	X124609	Target Value	.081
High Pressure	1100		
Alco Test Values	1st Alco	.074	2nd Alco
			.076
Signature			

(OVER)

DMB
6/23/11


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 2010.


 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JUNE 12, 2011

OPERATOR'S NAME: SMITH/PATRICK/J
OPERATOR'S NUMBER: 6397
SUBJECT'S LAST NAME: VOC
SUBJECT'S FIRST NAME/MI : VOC
O.L. #: 1234579
DEPT/AGENCY: CLM4
CASE/REPORT: 11-110611
TEST TYPE: V
ALCO TARGET VALUE: .001
ALCO S/N: X124609

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.29 in
ALCO TARGET .079 02:10
BLANK TEST .000 02:11
INTERNAL STANDARD VERIFIED 02:11
ALCO TV 29.29 in .074 02:11
BLANK TEST .000 02:12
SUBJECT SAMPLE .000 02:12
BLANK TEST .000 02:13
ALCO TV 29.29 in .076 02:13
BLANK TEST .000 02:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JUNE 12, 2011
TIME 02:07

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~`