

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130493</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>SMITH, PATRICK</u>	ID# <u>6397</u> Date <u>26 JUL 2012</u>
A	
Agency <u>CLEAR SECURITY FORCES</u>	Phone # <u>585-6313</u>
Instrument Location <u>BLDG 37, CLEAR AFS, AK</u>	
B	
Alco S/N <u>X124609</u>	Target Value <u>.081</u> High Pressure <u>1075</u>
Alco Test Values	<u>.074</u> <u>.076</u>
1st Alco	2 nd Alco
Signature <u>[Signature]</u>	<u>COB 8/7/12</u>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Nov, 2012.
NAK 11/2/12

[Signature] (Notary Seal Stamp)
~~Daniel L. Nicholas~~ Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office 11/2/12



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JULY 26, 2012

OPERATOR'S NAME:
SMITH/PATRICK/M
OPERATOR'S NUMBER: 6397
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567
DEPT/AGENCY: CLR4
CASE/REPORT: VOC/20120726
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X124609

COB
9/7/12

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.57 in
ALCO TARGET .080 10:19
BLANK TEST .000 10:20
INTERNAL STANDARD VERIFIED 10:21
ALCO TV 29.57 in .074 10:21
BLANK TEST .000 10:22
SUBJECT SAMPLE COB 9/7/12 .000 10:22
BLANK TEST .000 10:23
ALCO TV 29.57 in .076 10:23
BLANK TEST .000 10:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JULY 26, 2012
TIME 10:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.57 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~

