

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

Supervisor/Operator Performing the Verification Procedure:

Name Patrick Smith ID# 6397 Date 06 Apr 09

A Agency Clear AFS Security Forces Phone # 585-6293

Instrument Location Bldg 37 Clear Air Force Station

Alco S/N X124609 Target Value .081 High Pressure 1200

B

Alco Test Values .077 .077
 1st Alco 2nd Alco


Signature [Signature] CB
5/13/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 14 day of May 2009.
[Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

APRIL 06, 2009 ✓ *CS 3/26/09*

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: CLRA

CASE/REPORT: 1234567

TEST TYPE: V

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X124569 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.60 in		
ALCO TARGET	.080	10:36
BLANK TEST	.000	10:37
INTERNAL STANDARD	VERIFIED	10:37
ALCO TV 29.60 in	.077 ✓	10:37
BLANK TEST	.000	10:38
SUBJECT SAMPLE	.080 ✓	10:39
BLANK TEST	.000 ✓	10:40
ALCO TV 29.60 in	.077	10:40
BLANK TEST	.000	10:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

APRIL 06, 2009 ✓

TIME 10:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUvwxyz[\]^_`abcde fghi jk lmnop
pqrstuvwxyz{|}~