

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130493</u> ✓	
Name <u>SMITH, PATRICK J.</u>	ID# <u>6397</u> Date <u>05-21-08</u> ✓
A Agency <u>CLEAR AFS</u> ✓	Phone # <u>907-585-6313</u>
Instrument Location <u>BLDG 37</u>	
Alco S/N <u>78268</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>825 PSI</u>	
B Alco Test Values <u>.078</u> ✓ <u>.078</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u>	<u>CSB</u> <u>6/5/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick J Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/5/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

MAY 21, 2008

OPERATOR'S NAME:

SMITH, PATRICK J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: CLR4

CASE/REPORT: 00-0000

TEST TYPE: V

ALCO TARGET VALUE: 001

ALCO S/N: 78268

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.41 in
ALCO TARGET .079 00:26
BLANK TEST .000 00:27
INTERNAL STANDARD VERIFIED 00:27
ALCO TV 29.41 in .078 00:28
BLANK TEST .000 00:28
SUBJECT SAMPLE .000 00:29
BLANK TEST .000 00:30
ALCO TV 29.41 in .078 00:30
BLANK TEST .000 00:31

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

MAY 21, 2008

TIME 00:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 51c

BREATH TUBE: 41c

BAROMETER: 29.41 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~