

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

May emb
APR 13 2011

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130526</u>
A	Name	<u>JOSEPH V. SHUMAKER</u>	ID# <u>4188</u> Date <u>5-6-11</u>
	Agency	<u>SDP1 (SAND POINT P.D.)</u>	Phone # <u>907-383-3700</u>
	Instrument Location	<u>SAND POINT ALASKA 99661</u>	
B	Alco S/N	<u>3B</u>	Target Value <u>.080</u> High Pressure <u>1050</u>
	Alco Test Value Average	<u>.078</u>	<u>.076</u>
		1 st Alco	2 nd Alco
	Signature	<u>Joseph Shumaker</u>	

(OVER)

emb
5/13/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2011. NJB

Daniel L. Nicholas (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130526

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

MAY 06, 2011

OPERATOR'S NAME:

SUPERVISOR

OPERATOR'S NUMBER: 4188

SUBJECT'S LAST NAME:

SHOENAKER

SUBJECT'S FIRST NAME/MI :

J

O.L. #: VERIFICATION OF CAL

DEPT/AGENCY: SUP1

CASE/REPORT: L

TEST TYPE: V

ALCO TARGET VALUE: 000

ALCO S/N: ~~130526~~ 3B

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.07 in
ALCO TARGET .000 18:00
BLANK TEST .000 18:09
INTERNAL STANDARD VERIFIED 18:09
ALCO TV 30.07 in .078 18:09
BLANK TEST .000 19:10
SUBJECT SAMPLE ~~000~~ .000 18:11
BLANK TEST .000 18:12
ALCO TV 30.09 in .076 18:12
BLANK TEST .000 18:13

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

MAY 06, 2011

TIME 18:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 42c

BAROMETER: 30.09 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcedefghijklmnop
qrstuvwxyz{|}~`