

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130526 ✓
Name		Joseph V. SHOEMAKER	ID# 4188 Date 8-7-08 ✓
A	Agency	SBP1 SAND POINT P.D. ✓	Phone# 907-383-3700
Instrument Location SAND POINT, ALASKA 99669			
Alco S/N		77621 ✓	Target Value .084 ✓ High Pressure 875
B	Alco Test Value Average		
		.081	.081
		1 st Alco	2 nd Alco
Signature			Joseph Shoemaker
			8/18/08

(OVER)


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Joseph V Shoemaker, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

8/20/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130526 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

AUGUST 07, 2008

OPERATOR'S NAME:

SHOEMAKER JV

OPERATOR'S NUMBER: 4100

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

S

O.L. #: S

DEPT/AGENCY: SDP1

CASE/REPORT: K

TEST TYPE: V

ALCO TARGET VALUE: .004 ✓

ALCO S/N: 77621 ✓

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.27 in
ALCO TARGET .004 15:16
BLANK TEST .000 15:17
INTERNAL STANDARD VERIFIED 15:17
ALCO TV 30.27 in .001 15:18
BLANK TEST .000 15:19
SUBJECT SAMPLE .000 15:19
BLANK TEST .000 15:20
ALCO TV 30.27 in .001 15:20
BLANK TEST .000 15:22

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

AUGUST 07, 2008 ✓

TIME 15:23

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 06/26/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 30.27 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./:;<=>?`{|}~!@#&'()*+,-./:;<=>?`{|}~!
HIJKLMNOPQRSTUVWXYZ[]^_`abcde fghijklmno
pqrstuvwxyz{|}~!@#&'()*+,-./:;<=>?`{|}~!