

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130526 ✓
Name	JOSEPH V. SHOEMAKER	ID#	4188
		Date	10-5-08 ✓
A	Agency	SDPI SAND POINT P.D.	Phone # 907-383-3700
	Instrument Location	SAND POINT, AK 99661 ✓	
Alco S/N		77621 ✓	Target Value .084 ✓
			High Pressure 800 ✓
B	Alco Test Value Average	.081 ✓	.081 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		Joseph V. Shoemaker	
		CDS 10/24/08	

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Joseph V Shoemaker, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

10/24/08  
Date

