

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130526</u> ✓
Name	<u>Joseph V. Stoemaker</u>	ID#	<u>4188</u> ✓
Date	<u>11-15-10</u> ✓		
Agency	<u>(SDPI) SAND POINT</u>	Phone #	<u>907-383-3700</u>
Instrument Location	<u>SAND POINT, A.K 99661</u>		
Alco S/N	<u>Y173004</u> ✓	Target Value	<u>.079</u> ✓
High Pressure	<u>300</u>		
Alco Test Value Average	<u>.081</u> ^{.079} _{12/19/10} ^{DW}	<u>.079</u> ^{.080} _{12/19/10} ^{DW}	
	1 st Alco	2 nd Alco	
Signature	<u>Joseph V. Stoemaker</u>		

(OVER)

DMS
12/16/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of DEC, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130526

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

NOVEMBER 15, 2010

OPERATOR'S NAME:

SHOENAKER J V

OPERATOR'S NUMBER: 4188

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

H

O.L. #: K

DEPT/AGENCY: SDPI

CASE/REPORT: J

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: X173004

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.94 in
ALCO TARGET .081 15:39
BLANK TEST .000 15:41
INTERNAL STANDARD VERIFIED 15:41
ALCO TV 30.94 in .079 15:41
BLANK TEST .000 15:42
SUBJECT SAMPLE .000 15:43
BLANK TEST .000 15:43
ALCO TV 30.94 in .080 15:44
BLANK TEST .000 15:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

NOVEMBER 15, 2010

TIME 15:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 40c

BAROMETER: 30.94 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~