

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130567</u>	
Name <u>DEREK A. MCGARRIGAN</u>	ID# <u>5920</u> Date <u>5/25/08</u> ✓
Agency <u>KTN. POLICE DEPT. ✓</u>	Phone # <u>225-6631</u>
Instrument Location <u>Km. Police Dept. 361 MAIN ST. KETCHIKAN AK 99901</u>	
Alco S/N <u>X172938</u> ✓	Target Value <u>.076</u> ✓ High Pressure <u>800</u>
Alco-Test Value Average <u>.083</u> ✓	<u>.083</u> ✓
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>	<u>[Signature]</u> 6/5/08

(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Derek McGarrigan, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature] \_\_\_\_\_ Date 6/9/08  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130567

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130567

MAY 25, 2008

OPERATOR'S NAME:

MCBARRIGAN DEREK A

OPERATOR'S NUMBER: 5420

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI:

VOC

Q.L. #: 1234567

DEPT/AGENCY: KET1

CASE/REPORT: 08-1

TEST TYPE: V

ALCO. TARGET VALUE: .076

ALCO S/N: N172938

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 30.22 in		
ALCO TARGET	.076	07:55
BLANK TEST	.000	07:56
INTERNAL STANDARD	VERIFIED	07:56
ALCO TV 30.22 in	.003	07:56
BLANK TEST	.000	07:57
SUBJECT SAMPLE	.000	07:58
BLANK TEST	.000	07:59
ALCO TV 30.22 in	.003	07:59
BLANK TEST	.000	08:00

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130567

MAY 25, 2008

TIME 07:52

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	30.22 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ \* " # \$ % ' ( ) \* +