VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130567

Supervisor/Operator Performing the Verification Procedure:

A

Name Colleen O'Bryant ID# 6237 Date 6/11/10

Agency SCDL Phone # 269-5783

Instrument Location SCDL

B

Alco S/N Sco408 Target Value 0.080 High Pressure 450

Alco Test Values 0.084 (1st Alco) 0.083 (2nd Alco)

Signature C. O'Bryant CB 7/29/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 29th day of July, 2010.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130567

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130567

JUNE 11, 2010 CQB 7/29/10

OPERATOR'S NAME:
OBRYANT/COLLEEN
OPERATOR [illegible]: 6837
SUBJECT'S LAST NAME:
[illegible]
SUBJECT'S FIRST NAME/MI :
MOC
D.L. #: 0123456
DEPT/AGENCY: AAA1
CASE/REPORT: 10-1234
TEST TYPE: V
ALCO TARGET VALUE: 0.080
ALCO S/N: SCOTT80

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.51 in

ALCO TARGET	.078	11:38
BLANK TEST	.000	11:39
INTERNAL STANDARD	VERIFIED	11:39
ALCO TV 29.51 in	.084	11:40
BLANK TEST	.000	11:40
SUBJECT SAMPLE	.000	11:41
BLANK TEST	.000	11:42
ALCO TV 29.51 in	.083	11:42
BLANK TEST	.000	11:43

CQB 7/29/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130567

JUNE 11, 2010
TIME 11:39

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	40c
BAROMETER:	29.49 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde'ghijklmno
pqrstuvwxyz{|}~