

VERIFICATION OF CALIBRATION REPORT JUN 22 2009

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130567

Name DEREK A. MCGARREGAN ID# 5420 Date 6/13/09 ✓

A Agency Ketchikan Police Department Phone # 225-6631

Instrument Location 361 MAIN ST., KETCHIKAN AK 99901 ✓

---

B Alco S/N X 124662 Target Value .080 High Pressure 1200

Alco Test Value Average .084 ✓ .084 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature] ✓

CBS 6/25/09  
6/25/09

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of June, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 138567 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 138567

JUNE 13, 2009 ✓

OPERATOR'S NAME:

MCARRIGAN DEREK A

OPERATOR'S NUMBER: 5420

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: KET1

CASE/REPORT: 09-1

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X124662 ✓

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.09 in		
ALCO TARGET	.000	15:32
BLANK TEST	.000	15:33
INTERNAL STANDARD	VERIFIED	15:33
ALCO TV 30.09 in	.004 ✓	15:33
BLANK TEST	.000	15:34
SUBJECT SAMPLE	.000 ✓	15:34
BLANK TEST	.000 ✓	15:35
ALCO TV 30.09 in	.004 ✓	15:36
BLANK TEST	.000	15:36

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 138567

JUNE 13, 2009 ✓

TIME 15:26

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	30.09 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~