VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130567

Supervisor/Operator Performing the Verification Procedure:

A

Name DEREK A. McGARRIGAN ID# 5420 Date 7/31/08

Agency KETCHIKAN POLICE DEPARTMENT Phone # 225-6621

Instrument Location KTN. POLICE DEPT., 361 MAIN ST., KETCHIKAN AK. 99901

B

Alco S/N X172938 Target Value .076 High Pressure 400

Alco Test Value Average .082 (1st Alco) .083 (2nd Alco)

Signature

8/12/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Derek McGarrigan, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Signed and sealed,

8/13/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

AUG 0 1 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130567

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130567

JULY 31, 2008 8/12/08

OPERATOR'S NAME:
MCGARRIGAN DEREK A
OPERATOR'S NUMBER: 5420
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI.:
VOC
O.L. #: 123456
DEPT/AGENCY: KETI
CASE/REPORT: 08-111
TEST TYPE: V
ALCO TARGET VALUE: .076
ALCO S/N: X172938

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 30.14 in

ALCO TARGET	.076	06:47
BLANK TEST	.000	06:48
INTERNAL STANDARD	VERIFIED	06:48
ALCO TV 30.14 in	.082	06:49
BLANK TEST	.000	06:49
SUBJECT SAMPLE	.000	06:50
BLANK TEST	.000	06:51
ALCO TV 30.14 in	.083	06:51
BLANK TEST	.000	06:52

8/12/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130567

JULY 31, 2008
TIME 06:45

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	30.14 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~