

VERIFICATION OF CALIBRATION REPORT

MAR 15 2011

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130569 ✓
A	Name	Mac Burrell	ID # 4756 ✓ Date 3/10/2011 ✓
	Agency	Cordova Police Department	Phone # 424-6100
	Instrument Location	Cordova Police Department	
B	Alco S/N	65859 ✓	Target Value 0.081 ✓ High Pressure 1000
	Alco Test Value Average	0.081 ✓	0.081 ✓
		1st Alco	2nd Alco
	Signature	<i>Mac Burrell</i>	
		(OVER)	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1ST day of APRIL, 2011

Daniel L. Nicholas (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

DataMaster cdm S/N

130569

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130569

MARCH 10, 2011

OPERATOR'S NAME:
BURRELL/M

OPERATOR'S NUMBER: 4756

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

M

O.L. #: 1

DEPT/AGENCY: CDV1

CASE/REPORT: 031011

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: 65859

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.77 in
ALCO TARGET .000 01:19
BLANK TEST .000 01:20
INTERNAL STANDARD VERIFIED 01:20
ALCO TV 29.77 in .001 01:21
BLANK TEST .000 01:21
SUBJECT SAMPLE .000 01:23
BLANK TEST .000 01:24
ALCO TV 29.77 in .001 01:24
BLANK TEST .000 01:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130569

MARCH 10, 2011

TIME 01:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 29.77 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyzi!@#