


VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

AUG 21 2009

Supervisor/Operator Performing the Verification Procedure:	
Name <u>John Hodges</u>	ID # <u>2999</u> Date <u>8/19/2009</u>
Agency <u>Cordova Police Department</u>	Phone # <u>424-6100</u>
Instrument Location <u>Cordova Police Department</u>	
DataMaster cdm S/N <u>130569</u>	
Alco S/N <u>X124634</u>	Target Value <u>0.078</u> High Pressure <u>500</u>
Alco Test Value Average <u>0.081</u>	<u>0.081</u>
1st Alco	2nd Alco
Signature 	
(OVER) CJB 8/27/09	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory. (Do Not write in the area below)
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.




Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.


(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

DataMaster cdm S/N

130569 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130569

AUGUST 19, 2009 ✓ *CAB 8/17/09*

OPERATOR'S NAME:
HODGES, JOHN

OPERATOR'S NUMBER: 2999

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: CDV1

CASE/REPORT: X

TEST TYPE: V

ALCO TARGET VALUE: .078 ✓

ALCO S/N: K124604 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.20 in ✓
ALCO TARGET .078 05:35
BLANK TEST .000 05:36
INTERNAL STANDARD VERIFIED 05:36
ALCO TV 30.20 in .051 ✓ 05:36
BLANK TEST .000 05:37
SUBJECT SAMPLE .000 05:38
BLANK TEST .000 05:38
ALCO TV 30.20 in .001 ✓ 05:39
BLANK TEST .000 05:39

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130569

AUGUST 19, 2009 ✓

TIME 05:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.20 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~