

# VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

FEB 22 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130569
A	Name	MAC BURRELL	ID # 4756 Date 2/17/2010
	Agency	Cordova Police Department	Phone # 424-6100
	Instrument Location	Cordova Police Department	
B	Alco S/N	X301686	Target Value 0.086 High Pressure 850 PSI
	Alco Test Value Average	0.086	0.087
		1st Alco	2nd Alco
	Signature	<i>Mac Burrell</i>	
(OVER)			

*BMB*  
*2/23/2010*


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory. *(Do Not write in the area below)*
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program



Subscribed and sworn before me this 24th day of February, 2010.  
*Carolyn M. Noland* (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

DataMaster cdm S/N

130569

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130569

FEBRUARY 17, 2010

OPERATOR'S NAME:

BURRELL/M

OPERATOR'S NUMBER: 4756

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: CDVI

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .086

ALCO S/N: X301636

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 30.09 in

ALCO TARGET .086 19:40

BLANK TEST .000 19:41

INTERNAL STANDARD VERIFIED 19:41

ALCO TV 30.09 in .086 19:42

BLANK TEST .000 19:42

SUBJECT SAMPLE .087 19:43

BLANK TEST .000 19:44

ALCO TV 30.09 in .087 19:45

BLANK TEST .000 19:46

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130569

FEBRUARY 17, 2010

TIME 19:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 42c

BREATH TUBE: 42c

BAROMETER: 30.09 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKL/MNOPQRSTUVWXYZ\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~