

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 23 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130570</u> ✓
A	Name <u>C. RITALA</u>	ID# <u>3791</u> Date <u>4/21/10</u> ✓
	Agency <u>APD</u>	Phone # <u>786-2640</u>
	Instrument Location <u>FLORA SUB STATION (4th AVE)</u> ✓	
B	Alco S/N <u>78855</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>200</u>
	Alco Test Values	<u>.077</u> ✓ <u>.077</u> ✓ 1 st Alco 2 nd Alco
	Signature <u>[Handwritten Signature]</u> <u>RITALA/1349</u> ✓	
<u>CRB</u> <u>4/26/10</u>		

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 20th day of April, 2010.

[Handwritten Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

APRIL 21, 2010

OPERATOR'S NAME: RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :


A
O.L. #: A
DEPT/AGENCY: ANCI
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.77 in
ALCO TARGET .078 22:31
BLANK TEST .000 22:32
INTERNAL STANDARD VERIFIED 22:32
ALCO TV 29.77 in .077 22:32
BLANK TEST .000 22:33
SUBJECT SAMPLE .000 22:33
BLANK TEST .000 22:34
ALCO TV 29.77 in .077 22:34
BLANK TEST .000 22:35



DIAGNOSTIC CHECK



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

APRIL 21, 2010
TIME 22:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 51c
BREATH TUBE: 42c
BAROMETER: 29.77 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmnc
pqrstuvwxyz{|}~^

