

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 23 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130570 ✓

Name Derek Sitz ID# 4096 Date 4-23-10 ✓

A Agency Anchorage Police Department Phone # _____

Instrument Location Flora Substation

B Alco S/N X301600 ✓ Target Value .084 ✓ High Pressure 900

Alco Test Values .081 .080 ✓
1st Alco 2nd Alco

Signature [Signature] ✓

(OVER)

BMB
4/23/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 30th day of April, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

APRIL 22, 2010 ✓

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: ANCL

CASE/REPORT: 00-1234

TEST TYPE: V

ALCO TARGET VALUE: .084 ✓

ALCO S/N: X301600 ✓

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.81 in ✓

ALCO TARGET .083

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 29.81 in .081 ✓

BLANK TEST .000

SUBJECT SAMPLE .000 ✓

BLANK TEST .000 ✓

ALCO TV 29.81 in .080 ✓

BLANK TEST .000

BMB
4/28/10

BMB
4/28/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

APRIL 22, 2010 ✓

TIME 04:12

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 29.81 in

FLOW DETECTOR: OKAY

04:15 PUMP

04:16 HIGH SPEED: OKAY

04:16

04:16 DETECTOR: OKAY

04:17

04:17 FILTERS: OKAY

04:18

04:18 QUARTZ STANDARD: OKAY

04:19

CALIBRATION: OKAY

PRINTER TEST

! "\$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ Δ