

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 27 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130570
Name	Tadd McCouley	ID#	5996 5994
		Date	5/21/09
A	Agency	APD	Phone # 786-8763
	Instrument Location	4th Ave. Sub Station	✓
B	Alco S/N	XD2946 78895	5/27/09 5/27/09
	Target Value	.084	✓
	High Pressure	.1325	✓
	Alco Test Value Average	.080	✓
	1 st Alco	.080	✓
	2 nd Alco		
	Signature	<i>[Signature]</i>	

(OVER)

BMB
5/28/09


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of June 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

MAY 21, 2009

OPERATOR'S NAME:

MCCAULEY/TADD

OPERATOR'S NUMBER: 5994

SUBJECT'S LAST NAME:

MCCAULEY

SUBJECT'S FIRST NAME/MI :

TADD

O.L.L. #: 11111

DEPT/AGENCY: ANC1

CASE/REPORT: 0-000

TEST TYPE: T

ALCO TARGET VALUE: .004

ALCO S/N: 78895

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.27 in ✓
ALCO TARGET .004 ✓ 00:04
BLANK TEST .000 00:05
INTERNAL STANDARD VERIFIED 00:05
ALCO TV 30.27 in .000 00:05
BLANK TEST .000 00:06
SUBJECT SAMPLE .000 00:06
BLANK TEST .000 00:07
ALCO TV 30.27 in .000 00:07
BLANK TEST .000 00:08

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

MAY 21, 2009 ✓

TIME 00:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 51c
BREATH TUBE: 41c
BAROMETER: 30.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ\]^_`abcde+ghijklmno
pqrstuvwxyz{|}~