

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program JUL 09 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130570</u>
Name		<u>STEVE DUNN</u>	ID# <u>4106</u> Date <u>7-9-09</u>
A	Agency	<u>APD</u>	Phone # <u>786-8900</u>
Instrument Location		<u>4<sup>th</sup> AVE SUB.</u>	
Alco S/N		<u>78895</u>	Target Value <u>.084</u> High Pressure <u>500 PSI</u>
B	Alco Test Value Average		<u>.076</u> / <u>.077</u>
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<u>[Signature]</u> <u>7/14/09</u>	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2<sup>nd</sup> day of Jul, 2009.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office.

