VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

Supervisor/Operator Performing the Verification Procedure:

A

Name J Swartz ID# 4314 Date 08/06/08 ✓

Agency SCDL Phone # 269-5592

Instrument Location SCDL ✓

B

Alco S/N SC0TI07 ✓ Target Value .080 ✓ High Pressure 400 psi

Alco Test Values .077 ✓ (1st Alco) .078 ✓ (2nd Alco)

Signature [signature] CDB 8/18/08

(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do Not Write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 8/20/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

AUGUST 06, 2008 ✓ CJB 8/8/08

OPERATOR'S NAME:
SWARTZ/J
OPERATOR'S NUMBER: 4314
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: AAA1
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .080 ✓
ALCO S/N: SCOTT07 ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.96 in

ALCO TARGET	.080	09:35
BLANK TEST	.000	09:35
INTERNAL STANDARD	VERIFIED	09:36
ALCO TV 29.96 in	.077 ✓	09:36
BLANK TEST	.000	09:37
SUBJECT SAMPLE	.000 CJB 8/8/08	09:37
BLANK TEST	.000	09:38
ALCO TV 29.96 in	.078 ✓	09:38
BLANK TEST	.000	09:39

8/6/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

AUGUST 06, 2008 ✓
TIME 09:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	43c
BAROMETER:	29.96 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~