

AUG 19 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130570</u> ✓
Name	<u>Derek Sitz</u>	ID#	<u>4096</u> ✓ Date <u>8-18-10</u> ✓
A Agency	<u>Anchorage Police Department</u>	Phone #	<u>786-8900</u>
Instrument Location	<u>Flora Substation</u>		
B Alco S/N	<u>X301600</u> ✓	Target Value	<u>.084</u> ✓ High Pressure <u>300</u>
Alco Test Values	<u>.081</u> ✓ 1 st Alco	<u>.086</u> ✓ 2 nd Alco	
Signature	<u>Derek Sitz</u>		<u>003</u> <u>8/19/10</u>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of SEPT, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130570
CBS 8/19/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

AUGUST 18, 2010

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: ANCI

CASE/REPORT: 00-1234

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: W301600

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.91 in
ALCO TARGET .083 01:18
BLANK TEST .000 01:19
INTERNAL STANDARD VERIFIED 01:18
ALCO TV 29.90 in .001 01:19
BLANK TEST .000 01:19
SUBJECT SAMPLE .084 01:20
BLANK TEST .000 01:21
ALCO TV 29.91 in .080 01:21
BLANK TEST .000 01:22

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

AUGUST 18, 2010

TIME 01:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/99

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 29.90 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abode-ghijklmnop
qrstuvwxyz{|}~