

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 10 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130570</u>
Name	<u>Derek Sitz</u>	ID#	<u>409L</u> Date <u>9.10.09</u>
A Agency	<u>Anchorage Police Dept</u>	Phone #	<u>786.8900</u>
Instrument Location	<u>Flora Substation</u>		
B Alco S/N	<u>78895</u>	Target Value	<u>.084</u> High Pressure <u>500 psi</u>
Alco Test Values	<u>.079</u> 1 st Alco	<u>.078</u> 2 nd Alco	
Signature	<u>[Signature]</u>		

(OVER)

BMB
9/22/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Oct., 2009.

[Signature]
(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

SEPTEMBER 10, 2009

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: ANCI

CASE/REPORT: 00-1234

TEST TYPE: V

ALCO TARGET VALUE: .084 ✓

ALCO S/N: 78895

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.65 in
ALCO TARGET .083
BLANK TEST .000
INTERNAL STANDARD VERIFIED
ALCO TV 29.65 in .079 ✓
BLANK TEST .000 ✓
SUBJECT SAMPLE .000
BLANK TEST .000
ALCO TV 29.65 in .078 ✓
BLANK TEST .000

Smb
9/22/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

SEPTEMBER 10, 2009 ✓
TIME 02:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 29.64 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

DS
1465
9.10.09

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~Δ