

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130570 ✓
A	Name	C. RITALA	ID# 3791 Date 10/20/08 ✓
	Agency	APD	Phone # 786.2640
	Instrument Location	Floor Substation (4th Avenue)	✓
B	Alco S/N	X124611 ✓	Target Value .072 ✓ High Pressure 900 ✓
	Alco Test Values	076 ✓ 1 st Alco	076 ✓ 2 nd Alco
	Signature	Chris Ritala /1349 ✓	COB 11/8/08


(OVER)

Under the Alaska Rules of Evidence, I certify that;
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/19/08
Date

OCT 21 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130570 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

OCTOBER 20, 2008

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI:

NEW ALCO TANK

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: .072 ✓

ALCO S/N: X124611

--- BREATH ANALYSIS ---

.072 ADJUSTED FOR 29.62 in
ALCO TARGET .071 16:07
BLANK TEST .000 16:08
INTERNAL STANDARD VERIFIED 16:08
ALCO TV 29.62 in .076 ✓ 16:08
BLANK TEST .000 16:09
SUBJECT SAMPLE .000 16:09
BLANK TEST .000 16:10
ALCO TV 29.62 in .076 ✓ 16:11
BLANK TEST .000 16:11

cel/bcy 10/20/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

OCTOBER 20, 2008 ✓
TIME 16:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 29.62 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPQRSTUWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~

cel/bcy 10/20/08