VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

Supervisor/Operator Performing the Verification Procedure:

A

Name Glen Daily ID# 2528 Date 12/5/08

Agency Anchorage Police Phone # 786-2434

Instrument Location 4th Ave Sub

B

Alco S/N X301563 Target Value .080 High Pressure 1200

Alco Test Value Average ~~.079~~ .077 (1st Alco) 12-11-08 CAW ~~.076~~ 0.076 (2nd Alco) CDB 12/18/08

Signature CDB 12/23/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Glen Daily , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

12/23/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

DEC 11 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

DECEMBER 05, 2008 CAB 12/23/08

OPERATOR'S NAME:
DAILY/GLEN/P
OPERATOR'S NUMBER: 2528
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
TEST
O.L. #: 0000000
DEPT/AGENCY: ANC1
CASE/REPORT: TEST
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: X301563

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.73 in

ALCO TARGET	.079	02:36
BLANK TEST	.000	02:37
INTERNAL STANDARD	VERIFIED	02:37
ALCO TV 29.73 in	.077	02:38
BLANK TEST	.000	02:38
SUBJECT SAMPLE	000	02:39
BLANK TEST	.000	02:40
ALCO TV 29.73 in	.076	02:40
BLANK TEST	.000	02:41

CAB 12/23/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

DECEMBER 05, 2008
TIME 02:42

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	29.73 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~