

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130570 ✓
Name		C. RITALA	ID# 3791 Date 12/16/08 ✓
A	Agency	ANCHORAGE POLICE DEPT.	Phone # _____
Instrument Location		FLORA SUBSTATION (4th AVE)	✓
Alco S/N		X 301563 ✓	Target Value .080 ✓ High Pressure 900
Alco Test Value Average		.080 ✓	.079 ✓
		1 st Alco	2 nd Alco
Signature			RITALA/1349 ✓ COB 12/23/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/23/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

DECEMBER 16, 2008 ✓

OPERATOR'S NAME: *COB 12/16/08*
RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

A

O/L #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X301563

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.55 in
ALCO TARGET .001 14:28
BLANK TEST .000 14:29
INTERNAL STANDARD VERIFIED 14:29
ALCO TV 30.55 in .000 ✓ 14:30
BLANK TEST .000 14:30
SUBJECT SAMPLE *COB 12/16/08* .000 14:31
BLANK TEST .000 14:32
ALCO TV 30.55 in .079 ✓ 14:32
BLANK TEST .000 14:33

OK 1349 12/16/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

DECEMBER 16, 2008 ✓
TIME 14:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 30.55 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \ ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

OK 1349 12/16/08