

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130570 ✓
A	Name	C. RITALA	ID# 3791 Date 12/22/08 ✓
	Agency	APD	Phone # 786-2640
	Instrument Location	FLORA SUB STATION	✓
B	Alco S/N	X301563 ✓	Target Value .080 ✓ High Pressure 800
	Alco Test Value Average	.078	.078 ✓
		1 st Alco	2 nd Alco
	Signature		RITALA/1345 ✓ COR 12/30/08
(OVER)			

Under the Alaska Rules of Evidence, I certify that (write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/31/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument --

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570
DECEMBER 22, 2008
OPERATOR'S NAME: RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI.: A
O.L. #: A
DEPT/AGENCY: ANCI
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: X301563

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.03 in		
ALCO TARGET	.000	16:28
BLANK TEST	.000	16:28
INTERNAL STANDARD	VERIFIED	16:29
ALCO TV 30.03 in	.078	16:29
BLANK TEST	.000	16:30
SUBJECT SAMPLE	.000	16:30
BLANK TEST	.000	16:31
ALCO TV 30.03 in	.078	16:31
BLANK TEST	.000	16:32

1349 12/22/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570
DECEMBER 22, 2008
TIME 16:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	30.03 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy{|}~*"

1349 12/22/08