

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130570</u> ✓
Name	<u>C. RITA</u>	ID# <u>3791</u> ✓ Date <u>3/3/11</u> ✓
A Agency	<u>APD</u>	Phone # <u>786 2640</u>
Instrument Location <u>4th Ave 53 (Floor)</u>		
B Alco S/N	<u>X124611</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>700</u>
Alco Test Values	<u>.080</u> ✓ 1st Alco	<u>.080</u> ✓ 2 nd Alco
Signature	<u>[Signature]</u>	<u>RITA/1349</u>

(OVER)

BMB
3/6/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR, 2011 aw

[Signature]
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

MARCH 03, 2011

OPERATOR'S NAME:
RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/NI :
SIXTY DAY

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: .051

ALCO S/N: 3124511

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.77 in		
ALCO TARGET	.000	00:23
BLANK TEST	.000	00:24
INTERNAL STANDARD	VERIFIED	00:24
ALCO TV 29.77 in	.000	00:24
BLANK TEST	.000	00:25
SUBJECT SAMPLE	.000	00:25
BLANK TEST	.000	00:26
ALCO TV 29.75 in	.000	00:26
BLANK TEST	.000	00:27

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

MARCH 03, 2011
TIME 00:21

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/00
HEATERS	
SAMPLE CHAMBER:	50c
DEPTH TUBE:	41c
BAROMETER:	29.75 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY
PRINTER TEST	
!"#\$%&'()*+,-./0123456789:;<=>@ABCDEFGHIJKL MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstu vwxyz{ }~	

BMB
3/6/11

CR/1349

CR/1349

3/3/11

3/3/11