VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571

Supervisor/Operator Performing the Verification Procedure:

A

Name JSwartz ID# 4314 Date 3/10/08

Agency SCDL Phone # 269-5592

Instrument Location SCDL

B

Alco S/N 5C0TT07 ✓ Target Value .080 ✓ High Pressure 600psi

Alco Test Values .077 (1st Alco) .078 (2nd Alco)

Signature [signature]

CDB 3/17/08

(OVER)

(Do Not Write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 3/18/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORIES

MAR / 3 2008

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

MARCH 10, 2008 CB 3/17/08

OPERATOR'S NAME:
SWARTZ/J
OPERATOR'S NUMBER: 4314
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: AAA1
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .08
ALCO S/N: SCOTT07

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.68 in

ALCO TARGET	.079	10:08
BLANK TEST	.000	10:10
INTERNAL STANDARD	VERIFIED	10:10
ALCO TV 29.68 in	.077	10:10
BLANK TEST	.000	10:11
SUBJECT SAMPLE	.000	10:12
BLANK TEST	.000	10:12
ALCO TV 29.68 in	.079	10:13
BLANK TEST	.000	10:14

CB 3/17/08

3/10/08

3/10/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

MARCH 10, 2008
TIME 09:54

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	52c
BREATH TUBE:	42c
BAROMETER:	29.67 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;(=)?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~