

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 28 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130571</u>
Name	<u>RICK STEIDING</u>	ID#	<u>4189</u> Date <u>5-27-10</u>
Agency	<u>ANCHORAGE POLICE DEP.</u>	Phone #	<u>907 786-2669</u>
Instrument Location	<u>SEELEY SUB</u>		
Alco S/N	<u>ZB</u>	Target Value	<u>1080</u> High Pressure <u>1100 PSI</u>
Alco Test Values	<u>.081</u>	<u>.081</u>	
	1 st Alco	2 nd Alco	
Signature	<u>[Signature]</u>	<u>1501</u>	<u>0013</u> <u>6/7/10</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of June, 2010.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

R.S. 1501

J.C. 1501

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571 ✓

MAY 27, 2010 ✓

MAY 27, 2010 ✓
TIME 12:49

OPERATOR'S NAME: *COB 6/7/10*
STEIDING
OPERATOR'S NUMBER: 4189
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
V
O.L. #: 00000
DEPT/AGENCY: ANCI
CASE/REPORT: 000
TEST TYPE: V
ALCO TARGET VALUE: .000 ✓
ALCO S/N: 2B ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/08
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 30.27 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.27 in ✓
ALCO TARGET .000 12:51
BLANK TEST .000 12:52
INTERNAL STANDARD VERIFIED 12:52
ALCO TV 30.27 in .001 12:52
BLANK TEST .000 12:53
SUBJECT SAMPLE *COB 6/7/10* .000 12:54
BLANK TEST .000 12:55
ALCO TV 30.27 in .001 12:55
BLANK TEST .000 12:56

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~