

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130571
Name	STEVE DUNN	ID#	4106
Date	9-22-08		
Agency	APD	Phone #	786-8900
Instrument Location	ANCHORAGE JAIL		
Alco S/N	X 301755	Target Value	.025
High Pressure			400
Alco Test Value Average	.068	1 <sup>st</sup> Alco	.075
		2 <sup>nd</sup> Alco	
ALCO WAS NOT ON TESTED WHAT WAS LEFT IN HOSE. (SB)			
Signature			CSB 11/4/08
(OVER)			

Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Steve P Dunn, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

11/5/08  
Date

SEP 23 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130571 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

*S. Dunn #1361*

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130571

SEPTEMBER 22, 2008

OPERATOR'S NAME:  
DUNN/S  
OPERATOR'S NUMBER: 4106  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI:  
V

O.L. #: V  
DEPT/AGENCY: ANCI  
CASE/REPORT: V  
TEST TYPE: V  
ALCO TARGET VALUE: .075 ✓  
ALCO S/N: X301755 ✓

### --- BREATH ANALYSIS ---

.075 ADJUSTED FOR 30.20 in		
ALCO TARGET	.075	13:48
BLANK TEST	.000	13:48
INTERNAL STANDARD	VERIFIED	13:49
ALCO TV 30.20 in	.068 ✓	13:49
BLANK TEST	.000	13:50
SUBJECT SAMPLE	.068 ✓	13:50
BLANK TEST	.000	13:51
ALCO TV 30.22 in	.075 ✓	13:51
BLANK TEST	.000	13:52

## DIAGNOSTIC CHECK

*S. Dunn #1361*

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130571

SEPTEMBER 22, 2008 ✓  
TIME 13:45

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	30.22 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fgh i.jk lmnop  
qrs tuvwxyz{|}~?"