VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 0 1 2008

DataMaster cdm S/N 130571

Supervisor/Operator Performing the Verification Procedure:

A

Name C. RITALA ID# 3791 Date 9/30/08

Agency A.P.D. Phone # 786-2640

Instrument Location ANCHORAGE JAIL

B

Alco S/N X124646 Target Value .083 High Pressure 105.0

Alco Test Values .082 (1st Alco) .082 (2nd Alco)

Signature Chris Ritala RITALA/1349 11/20/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;
(Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

11/21/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

OCT 0 2 2008

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

```
ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

SEPTEMBER 30, 2008
OPERATOR'S NAME:
  RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
  VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
  A
O.L. #: A
DEPT/AGENCY: ANC1
CASE/REPORT: A
TEST TYPE: Y
ALCO TARGET VALUE: .083
ALCO S/N: X124646

      --- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.85 in
ALCO TARGET          .082     18:01
BLANK TEST           .000     18:02
INTERNAL STANDARD    VERIFIED 18:02
ALCO TV 29.85 in     .082     18:02
BLANK TEST           .000     18:03
SUBJECT SAMPLE       .000     18:04
BLANK TEST           .000     18:05
ALCO TV 29.85 in     .082     18:05
BLANK TEST           .000     18:06
```

CDB 11/20/08

CR/1349 9/30/08

DIAGNOSTIC CHECK

```
ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

SEPTEMBER 30, 2008
TIME 17:59

      --- DIAGNOSTIC CHECK ---

COMPUTER:            OKAY

PROGRAM:             OKAY

SOFTWARE DATE:       02/20/01

HEATERS
SAMPLE CHAMBER:      50c
BREATH TUBE:         41c

BAROMETER:           29.85 in

FLOW DETECTOR:       OKAY

PUMP
HIGH SPEED:          OKAY

DETECTOR:            OKAY

FILTERS:             OKAY

QUARTZ STANDARD:     OKAY

CALIBRATION:         OKAY

          PRINTER TEST
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~
```

CR/1349 9/30/08