

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130571 -</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>C. RITALA</u>	ID# <u>3791</u> Date <u>12/7/08 -</u>
A Agency <u>APD</u>	Phone # <u>786-2640</u>
Instrument Location <u>ANCHORAGE JAIL</u> ✓	
B Alco S/N <u>58427</u> ✓ Target Value <u>.076</u> ✓ High Pressure <u>1300</u>	
Alco Test Values <u>.075</u> ✓ <u>.075</u> ✓ 1 st Alco 2 nd Alco	
Signature <u></u> <u>RITALA/1349</u> ✓ <u>CRS</u> <u>12/18/08</u>	
(OVER)	


Under the Alaska Rules of Evidence, I certify that;
 (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



 12/20/08
 Jeanne Swartz Date
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N (30571)

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

DECEMBER 07, 2008

OPERATOR'S NAME:
RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/NI :
A
O.L. #: A
DEPT/AGENCY: ANCI
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .076 ✓
ALCO S/N: 50427

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 30.01 in		
ALCO TARGET	.076	03:10
BLANK TEST	.000	03:10
INTERNAL STANDARD	VERIFIED	03:10
ALCO TV 30.01 in	.075 -	03:11
BLANK TEST	.000	03:12
SUBJECT SAMPLE	.000	03:12
BLANK TEST	.000	03:13
ALCO TV 30.01 in	.075 -	03:14
BLANK TEST	.000	03:14

COB 12/18/08
CR/249 12/17/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

DECEMBER 07, 2008 ✓
TIME 03:07

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	41c
BAROMETER:	30.01 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

CR/249 12/17/08