

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

DECEMBER 10, 2010

OPERATOR'S NAME:
MARTIN, PATRICK E
OPERATOR'S NUMBER: 4072
SUBJECT'S LAST NAME:
MARTIN
SUBJECT'S FIRST NAME/MI :
P E

D.L. #: 1234567
DEPT/AGENCY: AHC1
CASE/REPORT: WOC
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: 2V

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.86 in

ALCO TARGET	.079	11:38
BLANK TEST	.000	11:38
INTERNAL STANDARD	VERIFIED	11:38
ALCO TV 29.86 in	.050	11:39
BLANK TEST	.000	11:39
SUBJECT SAMPLE	.000	11:39
BLANK TEST	.000	11:41
ALCO TV 29.86 in	.050	11:41
BLANK TEST	.000	11:42

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

DECEMBER 10, 2010
TIME 11:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/06
HEATERS
SAMPLE CHAMBER: 51c
BREATH TUBE: 42c
BAROMETER: 29.86 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HT,KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~*