

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130571 ✓
Name		C. RITALA	ID# 3791 Date 12/30/08 ✓
A	Agency	APD	Phone # 786-2640
Instrument Location		Anchorage Jail	✓
Alco S/N		70729 ✓	Target Value .076 ✓ High Pressure 1050
B	Alco Test Value Average		080 ✓ 080 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		Chris Ritala / 1349 ✓	008 1/6/09
(OVER)			

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

1/6/09  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130571

DECEMBER 30, 2008 ✓  
*CS 11/10/09*

OPERATOR'S NAME: RITALA  
OPERATOR'S NUMBER: 3791  
SUBJECT'S LAST NAME:  
VERIFICATION OF CALIBRATION  
SUBJECT'S FIRST NAME/NI: A  
O.L. #: A  
DEPT/AGENCY: ANCI  
CASE/REPORT: A  
TEST TYPE: V  
ALCO TARGET VALUE: .076 ✓  
ALCO S/N: 70729

### --- BREATH ANALYSIS ---

.076 ADJUSTED FOR 30.20 in		
ALCO TARGET	.076	16:53
BLANK TEST	.000	16:54
INTERNAL STANDARD	VERIFIED	16:54
ALCO TV 30.20 in	.000 ✓	16:54
BLANK TEST	.000	16:55
SUBJECT SAMPLE	.000	16:56
BLANK TEST	.000	16:57
ALCO TV 30.20 in	.000 ✓	16:57
BLANK TEST	.000	16:58

*CR/1349 12/30/08*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130571

DECEMBER 30, 2008 ✓  
TIME 16:50

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	30.20 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqr=stuvwxyz{|}~

*CR/1349 12/30/08*