

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130571</u> ✓
Name		<u>Derek Sitz</u>	ID# <u>4096</u> ✓ Date <u>3-2-11</u> ✓
A	Agency	<u>Anchorage Police Department</u>	Phone # <u>786-8900</u>
	Instrument Location	<u>Seely Substation</u>	
B	Alco S/N	<u>2V</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>700 psi</u>
	Alco Test Values	<u>.082</u> ✓ 1st Alco	<u>.082</u> ✓ 2nd Alco
	Signature	<u>[Signature]</u>	

(OVER)

BMB
3/6/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR, 2011 DW

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

MARCH 02, 2011

OPERATOR'S NAME:
SITZ, BEREK M
OPERATOR'S NUMBER: 4096
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 1234567
DEPT/AGENCY: ANCI
CASE/REPORT: 00-1234
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: 2V

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.24 in
ALCO TARGET .000 00:34
BLANK TEST .000 00:35
INTERNAL STANDARD VERIFIED 00:35
ALCO TV 30.24 in .002 00:35
BLANK TEST .000 00:36
SUBJECT SAMPLE .000 00:36
BLANK TEST .000 00:37
ALCO TV 30.24 in .000 00:37
BLANK TEST .000 00:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

MARCH 02, 2011
TIME 03:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/09 DS

HEATERS
SAMPLE CHAMBER: 50c DS
BREATH TUBE: 42c DS

BAROMETER: 30.24 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!*"#&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde(fghi)jklmno
pqrstuvwxyz{|}~*