

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 20 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130578

Name Michael A. Wery ID# 3090 Date 4-23-10

A Agency AST Phone # 457-2950

Instrument Location Cantwell

B Alco S/N X172969 Target Value .085 High Pressure 1200

Alco Test Value Average .076 .078
1st Alco 2nd Alco

Signature Michael A. Wery (OVER) CAB 5/24/10



I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (OVER)
I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory is an
in the Scientific Crime Detection Laboratory is an
in that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons
responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for
maintaining the records of the program.

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of June, 2010.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130578

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130578

APRIL 23, 2010

OPERATOR'S NAME:

WEYER/MICHAEL/A

OPERATOR'S NUMBER: 3590

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

OF CALIBRATION

O.L. #: 0

DEPT/AGENCY: CNT0

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .085

ALCO S/N: X172969

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 27.95 in
ALCO TARGET .079 12:54
BLANK TEST .000 12:55
INTERNAL STANDARD VERIFIED 12:55
ALCO TV 27.95 in .076 12:56
BLANK TEST .000 12:56
SUBJECT SAMPLE .000 12:57
BLANK TEST .000 12:58
ALCO TV 27.95 in .078 12:58
BLANK TEST .000 12:59

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130578

APRIL 23, 2010
TIME 12:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 27.95 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKL MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~*