

JAN 14 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130579</u> ✓
Name	<u>JEFFREY ECBIE</u>	ID#	<u>6623</u> ✓
Date	<u>1-14-11</u> ✓		
A Agency	<u>Anchorage Police Dept</u>	Phone #	<u>786-8900</u>
Instrument Location	<u>Anchorage Jail 1400 E 4th AVE</u>		
B Alco S/N	<u>1E</u> ✓	Target Value	<u>.080</u> ✓
High Pressure	<u>1000</u>		
Alco Test Values	<u>.081</u> ✓ 1 st Alco	<u>.081</u> ✓ 2 nd Alco	
Signature	<u>[Handwritten Signature]</u>		

(OVER)

8MB
2/8/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 28th day of FEB, 2010 AW

[Handwritten Signature] (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130579

BMB
2/8/11

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JANUARY 14, 2011

OPERATOR'S NAME:

ELBIE/JEFFREY/L

OPERATOR'S NUMBER: 6623

SUBJECT'S LAST NAME:

ELBIE

SUBJECT'S FIRST NAME/MI :

J/L

O.L. #: 000000

DEPT/AGENCY: ANC1

CASE/REPORT: 11-000

TEST TYPE: V

ALCO TARGET VALUE: .080

ALCO S/N: 1E

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.25 in

ALCO TARGET .080

BLANK TEST .000

INTERNAL STANDARD VERIFIED 02:38

ALCO TV 30.25 in .081

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 30.25 in .081

BLANK TEST .000

JE
30 lbs
1-14-11

BMB
2/8/11

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JANUARY 14, 2011

TIME 02:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 42c

BAROMETER: 30.25 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

JE
1-14-11
30 lbs

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~△