

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 21 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130579

Name Steve Dunn ID# 4106 Date 1-21-10


A Agency APD Phone # 764-1361

Instrument Location ANCHORAGE TALL

B Alco S/N 78895 Target Value .081 High Pressure 200 PSI

Alco Test Value Average .080 .084


1st Alco 2nd Alco

Signature 


(OVER) CB
1/22/10


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of January, 2010.

 (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JANUARY 21, 2010

OPERATOR'S NAME: *CS 1/22/10*
DUNN/S
OPERATOR'S NUMBER: 4106
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
V
O.L. #: V
DEPT/AGENCY: ANCL
CASE/REPORT: V
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: 78895

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.49 in		
ALCO TARGET	.079	04:09
BLANK TEST	.000	04:10
INTERNAL STANDARD	VERIFIED	04:10
ALCO TV 29.49 in	.080	04:10
BLANK TEST	.000	04:11
SUBJECT SAMPLE	.000	04:12
BLANK TEST	.000	04:13
ALCO TV 29.51 in	.084	04:13
BLANK TEST	.000	04:14

S. Dunn
A 1361

BT9 03/00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JANUARY 21, 2010
TIME 04:06

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.49 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~Δ

S. Dunn
A 1361