

VERIFICATION OF CALIBRATION REPORT MAR 10 2010
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130579</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Douglas Henie</u>	ID# <u>3554</u> Date <u>3-10-10</u> ✓
A Agency <u>APD</u>	Phone # <u>786-8500</u>
Instrument Location <u>Anchorage Jail</u> ✓	
B Alco S/N <u>718493</u> ✓ Target Value <u>.082</u> High Pressure <u>1000</u>	
Alco Test Values <u>.081</u> ✓ <u>.082</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>Douglas S Henie</u> ✓	
(OVER)	

SMB
3/24/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.

Carolyn M. Noland
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579 ✓

MARCH 10, 2010 ✓

OPERATOR'S NAME:

HENIE DOUGLAS S

OPERATOR'S NUMBER: 3554

SUBJECT'S LAST NAME:

TEST VOC

SUBJECT'S FIRST NAME/MI :

XXX

O.L. #: 111111

DEPT/AGENCY: ANCI

CASE/REPORT: 10-0000

TEST TYPE: D

ALCO TARGET VALUE: .082 ✓

ALCO S/N: 718493

BMB
3/24/10

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579 ✓

MARCH 10, 2010 ✓

TIME 04:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.36 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.36 in

ALCO TARGET .080

BLANK TEST .000

INTERNAL STANDARD VERIFIED 04:03

ALCO TV 29.36 in .081 ✓ 04:04

BLANK TEST .000 04:04

SUBJECT SAMPLE .000 04:05

BLANK TEST .000 04:05

ALCO TV 29.38 in .082 ✓ 04:06

BLANK TEST .000 04:07

BLANK TEST .000 04:07

BLANK TEST .000 04:08

BMB
3/24/10

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~Δ