

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579 ✓

Supervisor/Operator Performing the Verification Procedure:

Name Derek S. Iz ID# 4096 Date 4-11-09 ✓

A Agency Anchorage Police Dept Phone # _____

Instrument Location Anchorage Ia.1 ✓

B Alco S/N X 124630 ✓ Target Value .084 ✓ .082 ✓ 4-23-09 CDM High Pressure 450

Alco Test Values .081 ✓ 1st Alco .081 ✓ 2nd Alco

Signature Derek S. Iz ✓ OB
5/13/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz

Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14 day of May, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

[Empty box for taping the test strip]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

APRIL 11, 2009 ✓

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: ANC1

CASE/REPORT: 00-1234

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: X124630 ✓

DS
1465
9-11-09

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.47 in
ALCO TARGET .082 01:46
BLANK TEST .000 01:47
INTERNAL STANDARD VERIFIED 01:47
ALCO TV 29.47 in .081 ✓ 01:47
BLANK TEST .000 01:48
SUBJECT SAMPLE .000 01:49
BLANK TEST .000 01:50
ALCO TV 29.47 in .081 ✓ 01:50
BLANK TEST .000 01:51

DS
5/13/09

[Empty box for taping the test strip]

DIAGNOSTIC CHECK

[Empty box for taping the diagnostic check]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

APRIL 11, 2009 ✓

TIME 01:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY DS
PROGRAM: OKAY 1465
SOFTWARE DATE: 12/11/08 4-11-09
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 39c
BAROMETER: 29.47 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~`Δ

[Empty box for taping the diagnostic check]